

## Dr. Ram Manohar Lohia Hospital New Delhi

### Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )

Department : Trauma O.T. 5<sup>th</sup> Floor ( Surgery )

**Dated :- 04/04/15**

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
1.	Bench Top Autoclave S.N.-00/ R081593	“Andromeda Vacume XP”	Vishal Distributor 9811580946, 9212489578	Rd. from store 10/6/08	EURO 2,625/-	Monthly	28/01/14. by Mr. R.P. Verma Supervis by Sr. Tech. &Staff	Automatic water pump system not working,	Work on process from Workshop
2.	High Speed Autoclave Nat Steel S.N. 1715 (new small )	Nat Steel Equipment Pvt. Ltd.	Nat Steel Equipment Pvt. Ltd. 9971874796	Rd. from store 06/05/08	Cost Rs. 1,65,000/-	Monthly	26/02/15 by Mr. Opin  Supervis by Sr. Tech. &Staff	Functional	-----

3.	High Speed Autoclave S.N. 517 (new big )	Nat Steel	Nat Steel Equipments Ltd. 9818148715	Rd. from store 11/06/11	Cost Rs. 17,92,000/-	Monthly	24/03/15 by Mr. Raja Ram  Supervis by Sr. Tech. &Staff	Functional	-----
----	---	-----------	---	----------------------------	-------------------------	---------	--	------------	-------

## Dr. Ram Manohar Lohia Hospital New Delhi

### Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )

Department : Trauma O.T. 5<sup>th</sup> Floor( Surgery )

**Dated :- 04/04/15**

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
4.	Bowa Cautery a) S.N. 3500/336 b) S.N. 3500/681	Bowa Germany	Life Aid Device 9968370861	Tr. taken from E.O.T. 17/6/08 (681) &17/9/08 (336)	EURO 81,611/- each	Monthly	a) S.N. 3500/336 (04/03/2014) by Daleep from Rohit Surgical	S.N. 3500/336 Equipment is working but neutral electrode socket need to be changed	Cautery Machine is working  Work is on process



5.	Ceiling Light ( Three )	Berch told GMBH Germany	Rohit Surgical Pvt. Ltd. 9311088110	Rd. from store 06/05/08	EURO 19,396.5/- per unit	Monthly	24/03/15 by Mr. Daleep from Rohti Surgical  Supervis by Sr. Tech. &Staff	(Neuro O.T.) One Dome not working due to faulty control and power supply	Work on process from M &R
6.	O.T. Table Surgery T.S. 200 Takuchi S.N. 0080	Takuchi	Rohanika Elec. &Med. System 9212791757	Rd. from store 10/06/08	Yen 22,61,070/-	Monthly	07/03/15 by Mr. Bhavesh  Supervis by Sr. Tech. &Staff	Table is working  Only Right leg plate locking lever is not working properly needs to be replaced Work on process	-----
7.	O.T. Table Surgery (Surginox) S.no. 15068	Steris Surginox USA	Steris India Pvt. Ltd.	Rd. on 16/05/2012	\$ 15220/-	Monthly	17/12/14 by Mr. Amamul  Supervis by Sr. Tech. &Staff	Functional	-----

**Dr. Ram Manohar Lohia Hospital  
New Delhi**

**Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )**

Department : Trauma O.T. 5<sup>th</sup> Floor( Surgery )

**Dated :- 04/04/15**

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
8.	Ventilator 1). S.N. 3121118350 2). S.N. 3121123870 3). S.N. 3121124100 4). S.N. 3121124370	Covidien	Medi Works A-56, IIIrd floor, Lajpat Nagar-II, New Delhi. Ph : 91-11- 46563900, 93105 66668	Rd. from store 12/08/2013	Cost @ U.S. \$ 19,200/- each	Monthly	28/02/15 by Mr. Kumar Keshav  Supervis by Sr. Tech. &Staff	Functional	-----
9.	Post op. ICU Beds ( 04 ) 1. S.no. L231 HE 3606 2. S.no. M118 HE 1414 3. S.no. M113 HE 1388 4. S.no. M113 HE 1387  Model – 405 Electric Bed	Hill Rom	Rohanika Electronics &Medical Systems, New Delhi	Rd. on 05/01/2012	Rs. 1,47,232/-	Monthly	20/06/2014 by Mr. Ashok Kumar Sagar  Supervised by Sr. Tech. &Staff	Functional	-----

## Dr. Ram Manohar Lohia Hospital New Delhi

### Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )

Department : Trauma O.T. 5<sup>th</sup> Floor ( Anaesthesia )

**Dated :- 04/04/15**

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
1.	Defibrillator with AED S.N. US00544705 S.N. US00544706	Philips Heartstart Mrx	Suprmed Medical Equipments Pvt. Ltd. 9311001765	Rd. from store 13/10/10	Cost 4,30,000/- per unit	Monthly	26/03/2015 by Mr. Pramod  Supervised by Sr. Tech. & Staff	Functional	S.N. US00544705 SPO2 Sensor faulty need to be replaced

2.	<p>A). New Drager Anas. Machine Primus  a) S.N. ARZ0205  b) S.N. ARZ0206  c) S.N. ARZ0207  ( Tr. to F.F.O.T. on 16/11/11 )  with two Anaesthesia workstation with one compatible Central Monitoring Station</p>	<p>Drager Medical AG &amp;Co. K.G.  Germany  Moislinger Allee</p>	<p>H.L. Medical Systems Pvt. Ltd.  9810775642</p>	<p>Rd. from store  24/11/08</p>	<p>A). Euro 1,22,720/- each</p>	<p>Monthly</p>	<p>13/03/2015 for S.N. ARZ0205 &amp;09/03/2015 for S.N. ARZ0206 by Mr. Vikash</p> <p>Supervised by Sr. Tech. &amp;Staff</p>	<p>Functional</p> <p>S.N. ARZ0205 ECG 5 lead is not faulty, Estimate to be given by Com. Engineer, Machine is Working</p>	<p>-----</p>
----	---	---	---	-------------------------------------	---------------------------------	----------------	---	---	--------------

	B). Aneasthesia Work Station with Compatible Central Monitoring Station S.No. 10668 M 00836	Drager Medical AG &Co. K.G. Germany Moisinger Allee	H.L. Medical Systems Pvt. Ltd. 9810775642	Rd. from store 24/11/08	B). Euro 1,16,584/-	Monthly	02/03/2015 by Mr. Vishnoi  Supervised by Sr. Tech. &Staff	Functional  Mouse is not working, need to be replaced	-----
--	---	---	---	-------------------------	---------------------	---------	---	---	-------



## Dr. Ram Manohar Lohia Hospital New Delhi

### Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )

Department : Trauma O.T. 5<sup>th</sup> Floor ( Anaesthesia )

**Dated :- 04/04/15**

S. No.	Equipment Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
3.	Monitor Philips Intellivue MP 60M 8005A 1. S.N. DE73112724  2. S.N. DE73112728  3. S.N. DE73112730  4. S.N. DE73112731  5. S.N.	Philips	Philips Electronics India Ltd. New Delhi. 9311001765	Rd. from store 13/03/09	Cost \$ 2,95,000/-	Monthly	09/03/15 by Mr. Pramod  Supervised by Sr. Tech. & Staff	Functional	-----



4.	Defibrillator S.N. 01364	Cardiolife Nihon Kohden Corporation	Trivitrans Health care Pvt. Ltd. 9871800160	Rd. from Trauma ICU to T.O.T. On 08/9/08	Cost Yen 6,85,000/-	Monthly	22/05/13 by Mr. Dand Ali  Supervised by Sr. Tech. &Staff	Functional	-----
5.	Radio Frequency Generator  S. No. G-4-949	G4 Version 2 Cosman Medical Inc.	I-Meditools	Rd. on 14/10/14 DOI – 15/10/14	Rs. 29,85,000/-	Monthly	Demo given on 15/10/14	Functional	-----
6.	Transport Monitor  S. No. 140515014	-----	Allier Medicare	Rd. on 14/10/14 DOI – 17/10/14	Rs. @1,39,777/-	Monthly	Demo given on 15/10/14	Functional	-----

**Dr. Ram Manohar Lohia Hospital  
New Delhi**

**Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )**

Department : Trauma O.T. 5<sup>th</sup> Floor ( **FOR ORTHO** )

**Dated :- 04/04/15**

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
1.	O.T. Table Ortho S.N. 0806/00310	Zywies	Euro Meditools pvt. ltd. 9212401074	Rd. from Store 29/03/08	Euro 73,671/-	Monthly	01/04/15 by Mr. Bhola Supervised by Sr. Tech. &Staff	Functional	<b>Manual and Remote is not working for up down position since 03/11/14,</b>  <b>01/04/2015 O.T. Table working by direct online power source</b>
2.	Arthroscope S.N. 73951114	Smith &Nephew	Surgical System 995827440	Rd. from Store to F.F.O.T. 21/7/2006 Tr. Rd. From F.F.O.T. 8/11/08	Rs. 18,90,000/-	Monthly	09/08/2014 by Mr. Asim Supervised by Sr. Tech. &Staff	Functional	-----

## Dr. Ram Manohar Lohia Hospital New Delhi

### Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )

Department : Trauma O.T. 5<sup>th</sup> Floor ( **FOR ORTHO** )

**Dated :- 04/04/15**

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommened Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
3.	High Definition Image Recording system of Arthroscopy S.N. 4507 (72200242)	Smith &Nephew	Surgical System 995827440	Rd. from Store 27/10/2010	Rs. 7,25,000/-	Monthly	09/07/2013 by Mr. Asim  Supervised by Sr. Tech. &Staff,	Functional	-----

**Dr. Ram Manohar Lohia Hospital  
New Delhi**

**Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh ( Purchased on or after 01.04.2007 )**

Department : Trauma O.T. 5<sup>th</sup> Floor ( FOR ORTHO )

**Dated :- 04/04/15**

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
4.	Electric Drill & Saw S.N. 188035, 188036	Sodam Switzerland	Medi Link Surgical 9811040960	Rd. from store to FFOT 6/12/07 Tr. Rd. From FFOT to T.O.T. 16/9/2008	Rs. 9,66,500/-	Monthly	condemned	Not Functional	Condemned