

Dr. Ram Manohar Lohia Hospital New Delhi

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
1.	Bench Top Autoclave S.N.-00/ R081593	“Andromeda Vacume XP”	Vishal Distributor 9811580946, 9212489578	Rd. from store 10/6/08	EURO 2,625/-	Monthly	28/01/14. by Mr. R.P. Verma Supervis by Sr. Tech. &Staff	Automatic water pump system not working,	Work on process from Workshop
2.	High Speed Autoclave Nat Steel S.N. 1715 (new small)	Nat Steel Equipment Pvt. Ltd.	Nat Steel Equipment Pvt. Ltd. 9971874796	Rd. from store 06/05/08	Cost Rs. 1,65,000/-	Monthly	02/01/15 by Mr. Opin Supervis by Sr. Tech. &Staff	Functional	-----

3.	High Speed Autoclave S.N. 517 (new big)	Nat Steel	Nat Steel Equipments Ltd. 9818148715	Rd. from store 11/06/11	Cost Rs. 17,92,000/-	Monthly	24/12/14 by Mr. Raja Ram Supervis by Sr. Tech. & Staff	Functional	-----
----	---	-----------	---	----------------------------	-------------------------	---------	---	------------	-------

Dr. Ram Manohar Lohia Hospital New Delhi

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
4.	Bowa Cautery a) S.N. 3500/336 b) S.N. 3500/681	Bowa Germany	Life Aid Device 9968370861	Tr. taken from E.O.T. 17/6/08 (681) &17/9/08 (336)	EURO 81,611/- each	Monthly	a) S.N. 3500/336 (04/03/2014) by Daleep from Rohit Surgical b) S.N. 3500/681	S.N. 3500/336 Equipment is working but neutral electrode socket need to be changed S.N. 3500/681	Cautery Machine is working Work is on process

							(26/05/14) by Daleep from Rohit Surgical Supervis by Mr. Tech. &Staff	hand activation is not working	
--	--	--	--	--	--	--	--	--------------------------------	--

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
--------	--------------------	-------------------	------------------	------------------	---------------------	---	---	---------------------------	---

5.	Ceiling Light (Three)	Berch told GMBH Germany	Rohit Surgical Pvt. Ltd. 9311088110	Rd. from store 06/05/08	EURO 19,396.5/- per unit	Monthly	03/06/14 by Mr. Daleep from Rohti Surgical Supervis by Sr. Tech. &Staff	(Neuro O.T.) One Dome not working due to faulty control and power supply	Work on process from M &R
6.	O.T. Table Surgery T.S. 200 Takuchi S.N. 0080	Takuchi	Rohanika Elec. &Med. System 9212791757	Rd. from store 10/06/08	Yen 22,61,070/-	Monthly	10/09/14 by Mr. Bhavesh Supervis by Sr. Tech. &Staff	Functional	-----
7.	O.T. Table Surgery (Surginox) S.no. 15068	Steris Surginox USA	Steris India Pvt. Ltd.	Rd. on 16/05/2012	\$ 15220/-	Monthly	17/12/14 by Mr. Amamul Supervis by Sr. Tech. &Staff	Functional	-----

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
8.	Ventilator 1). S.N. 3121118350 2). S.N. 3121123870 3). S.N. 3121124100 4). S.N. 3121124370	Covidien	Medi Works A-56, IIIrd floor, Lajpat Nagar-II, New Delhi. Ph : 91-11- 46563900, 93105 66668	Rd. from store 12/08/2013	Cost @ U.S. \$ 19,200/- each	Monthly	23/12/14 by Mr. Kumar Keshav Supervis by Sr. Tech. &Staff	Functional	-----

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
9.	Post op. ICU Beds (04) 1. S.no. L231 HE 3606 2. S.no. M118 HE 1414 3. S.no. M113 HE 1388 4. S.no. M113 HE 1387 Model – 405 Electric Bed	Hill Rom	Rohanika Electronics &Medical Systems, New Delhi	Rd. on 05/01/2012	Rs. 1,47,232/-	Monthly	20/06/2014 by Mr. Ashok Kumar Sagar Supervised by Sr. Tech. &Staff	Functional	-----

Dr. Ram Manohar Lohia Hospital New Delhi

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (Anaesthesia)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
1.	Defibrilator with AED S.N. US00544705 S.N. US00544706	Philips Heartstart Mrx	Supramed Medical Equipments Pvt. Ltd. 9311001765	Rd. from store 13/10/10	Cost 4,30,000/- per unit	Monthly	03/12/2014 by Mr. Pramod Supervised by Sr. Tech. & Staff	Functional	----- S.N. US00544705 check defeb. Found charging problem due to battery faulty needs to be replaced

2.	<p>A). New Drager Anas. Machine Primus a) S.N. ARZ0205 b) S.N. ARZ0206 c) S.N. ARZ0207 (Tr. to F.F.O.T. on 16/11/11) with two Anaesthesia workstation with one compatible Central Monitoring Station</p>	<p>Drager Medical AG &Co. K.G. Germany Moislinger Allee</p>	<p>H.L. Medical Systems Pvt. Ltd. 9810775642</p>	<p>Rd. from store 24/11/08</p>	<p>A). Euro 1,22,720/- each</p>	<p>Monthly</p>	<p>24/12/2014 for S.N. ARZ0205 &24/12/2014 for S.N. ARZ0206 by Mr. Pinkul Supervised by Sr. Tech. &Staff</p>	<p>Functional</p>	<p>-----</p>
----	---	---	---	-------------------------------------	---------------------------------	----------------	--	-------------------	--------------

	B). Aneasthesia Work Station with Compatible Central Monitoring Station S.No. 10668 M 00836	Drager Medical AG &Co. K.G. Germany Moislinger Allee	H.L. Medical Systems Pvt. Ltd. 9810775642	Rd. from store 24/11/08	B). Euro 1,16,584/-	Monthly	24/12/2014 by Mr. Vishnoi Supervised by Sr. Tech. &Staff	Functional	-----
--	---	--	---	-------------------------	---------------------	---------	---	------------	-------

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (Anaesthesia)

Dated :- 07/01/15

S. No.	Equipment Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
3.	Monitor Philips Intellivue MP 60M 8005A 1. S.N. DE73112724 2. S.N. DE73112728 3. S.N. DE73112730 4. S.N. DE73112731 5. S.N. DE73112732 with trasport monitor S.N. 72813540 (One) Post Op. T.O.T. 5 th Floor	Philips	Philips Electronics India Ltd. New Delhi. 9311001765	Rd. from store 13/03/09	Cost \$ 2,95,000/-	Monthly	19/11/14 by Mr. Pramod Supervised by Sr. Tech. &Staff	Functional only NIBP cuffed adult and ECG trunk cable not working, Bed No. 3, S.N. DE73112730 Estimate provided on 12/12/2013, sent to Workshop with repair performa and propreitory certificate	work on process from workshop

Dr. Ram Manohar Lohia Hospital New Delhi

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (Anaesthesia)

Dated :- 07/01/15

S. No.	Equipment Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
4.	Defibrillator S.N. 01364	Cardiolife Nihon Kohden Corporation	Trivitrans Health care Pvt. Ltd. 9871800160	Rd. from Trauma ICU to T.O.T. On 08/9/08	Cost Yen 6,85,000/-	Monthly	22/05/13 by Mr. Dand Ali Supervised by Sr. Tech. & Staff	Functional	-----
5.	Radio Frequency Generator S. No. G-4-949	G4 Version 2 Cosman Medical Inc.	I-Meditools	Rd. on 14/10/14 DOI – 15/10/14	Rs. 29,85,000/-	Monthly	Demo given on 15/10/14	Functional	-----

6.	Transport Monitor S. No. 140515014	-----	Allier Medicare	Rd. on 14/10/14 DOI – 17/10/14	Rs. @1,39,777/-	Monthly	Demo given on 15/10/14	Functional	-----
----	--	-------	-----------------	---	--------------------	---------	---------------------------	------------	-------

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (FOR ORTHO)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising & Staff	Whether Functional or Not	If Not Functional details of Action taken
1.	O.T. Table Ortho S.N. 0806/00310	Zywies	Euro Meditools pvt. ltd. 9212401074	Rd. from Store 29/03/08	Euro 73,671/-	Monthly	06/11/14 by Mr. Bhola Supervised by Sr. Tech. & Staff	Not Functional	Manual and Remote is not working for up down position since 03/11/14, information given to M & R and Mr. Bhola

2.	Arthroscope S.N. 73951114	Smith &Nephew	Surgical System 995827440	Rd. from Store to F.F.O.T. 21/7/2006 Tr. Rd. From F.F.O.T. 8/11/08	Rs. 18,90,000/-	Monthly	09/08/2014 by Mr. Asim Supervised by Sr. Tech. &Staff	Functional	-----
----	------------------------------	---------------	------------------------------	--	--------------------	---------	---	------------	-------

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (FOR ORTHO)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
--------	--------------------	-------------------	------------------	------------------	---------------------	---	---	---------------------------	---

3.	High Definition Image Recording system of Arthroscopy S.N. 4507 (72200242)	Smith &Nephew	Surgical System 995827440	Rd. from Store 27/10/2010	Rs. 7,25,000/-	Monthly	09/07/2013 by Mr. Asim Supervised by Sr. Tech. &Staff,	Functional	-----
----	--	---------------	---------------------------	---------------------------	----------------	---------	---	------------	-------

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (FOR ORTHO)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
4.	Electric Drill &Saw S.N. 188035, 188036	Sodam Switzerland	Medi Link Surgical 9811040960	Rd. from store to FFOT 6/12/07 Tr. Rd. From FFOT to T.O.T. 16/9/2008	Rs. 9,66,500/-	Monthly	Condemn	Not Functional	Condemn

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
1.	Bench Top Autoclave S.N.- 00/ R081593	“Andromeda Vacume XP”	Vishal Distributor 9811580946, 9212489578	Rd. from store 10/6/08	EURO 2,625/-	Monthly	28/01/14. by Mr. R.P. Verma Supervis by Sr. Tech. &Staff	Automatic water pump system not working,	Work on process from Workshop

2.	High Speed Autoclave Nat Steel S.N. 1715 (new small)	Nat Steel Equipment Pvt. Ltd.	Nat Steel Equipment Pvt. Ltd. 9971874796	Rd. from store 06/05/08	Cost Rs. 1,65,000/-	Monthly	02/01/15 by Mr. Opin Supervis by Sr. Tech. &Staff	Functional	-----
3.	High Speed Autoclave S.N. 517 (new big)	Nat Steel	Nat Steel Equipments Ltd. 9818148715	Rd. from store 11/06/11	Cost Rs. 17,92,000/-	Monthly	24/12/14 by Mr. Raja Ram Supervis by Sr. Tech. &Staff	Functional	-----

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
--------	--------------------	-------------------	------------------	------------------	---------------------	---	---	---------------------------	---

4.	Bowa Cautery a) S.N. 3500/336 b) S.N. 3500/681	Bowa Germany	Life Aid Device 9968370861	Tr. taken from E.O.T. 17/6/08 (681) & 17/9/08 (336)	EURO 81,611/- each	Monthly	a) S.N. 3500/336 (04/03/2014) by Daleep from Rohit Surgical b) S.N. 3500/681 (26/05/14) by Daleep from Rohit Surgical Supervis by Mr. Tech. & Staff	S.N. 3500/336 Equipment is working but neutral electrode socket need to be changed S.N. 3500/681 hand activation is not working	Cautery Machine is working Work is on process
----	--	--------------	-------------------------------	--	--------------------	---------	---	--	--

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
5.	Ceiling Light (Three)	Berch told GMBH Germany	Rohit Surgical Pvt. Ltd. 9311088110	Rd. from store 06/05/08	EURO 19,396.5/- per unit	Monthly	03/06/14 by Mr. Daleep from Rohti Surgical Supervis by Sr. Tech. &Staff	(Neuro O.T.) One Dome not working due to faulty control and power supply	Work on process from M &R
6.	O.T. Table Surgery T.S. 200 Takuchi S.N. 0080	Takuchi	Rohanika Elec. &Med. System 9212791757	Rd. from store 10/06/08	Yen 22,61,070/-	Monthly	10/09/14 by Mr. Bhavesh Supervis by Sr. Tech. &Staff	Functional	-----
7.	O.T. Table Surgery (Surginox) S.no. 15068	Steris Surginox USA	Steris India Pvt. Ltd.	Rd. on 16/05/2012	\$ 15220/-	Monthly	17/12/14 by Mr. Amamul Supervis by Sr. Tech. &Staff	Functional	-----

Dr. Ram Manohar Lohia Hospital

New Delhi

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
8.	Ventilator 1). S.N. 3121118350 2). S.N. 3121123870 3). S.N. 3121124100 4). S.N. 3121124370	Covidien	Medi Works A-56, IIIrd floor, Lajpat Nagar-II, New Delhi. Ph : 91-11- 46563900, 93105 66668	Rd. from store 12/08/2013	Cost @ U.S. \$ 19,200/- each	Monthly	23/12/14 by Mr. Kumar Keshav Supervis by Sr. Tech. &Staff	Functional	-----

Dr. Ram Manohar Lohia Hospital

New Delhi

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 07/01/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
9.	Post op. ICU Beds (04) 1. S.no. L231 HE 3606 2. S.no. M118 HE 1414 3. S.no. M113 HE 1388 4. S.no. M113 HE 1387 Model – 405 Electric Bed	Hill Rom	Rohanika Electronics & Medical Systems, New Delhi	Rd. on 05/01/2012	Rs. 1,47,232/-	Monthly	20/06/2014 by Mr. Ashok Kumar Sagar Supervised by Sr. Tech. & Staff	Functional	-----

