

Dr. Ram Manohar Lohia Hospital New Delhi

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor (Surgery)

Dated :- 04/02/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
1.	Bench Top Autoclave S.N.-00/ R081593	“Andromeda Vacume XP”	Vishal Distributor 9811580946, 9212489578	Rd. from store 10/6/08	EURO 2,625/-	Monthly	28/01/14. by Mr. R.P. Verma Supervis by Sr. Tech. &Staff	Automatic water pump system not working,	Work on process from Workshop
2.	High Speed Autoclave Nat Steel S.N. 1715 (new small)	Nat Steel Equipment Pvt. Ltd.	Nat Steel Equipment Pvt. Ltd. 9971874796	Rd. from store 06/05/08	Cost Rs. 1,65,000/-	Monthly	02/01/15 by Mr. Opin Supervis by Sr. Tech. &Staff	Functional	-----

3.	High Speed Autoclave S.N. 517 (new big)	Nat Steel	Nat Steel Equipments Ltd. 9818148715	Rd. from store 11/06/11	Cost Rs. 17,92,000/-	Monthly	24/12/14 by Mr. Raja Ram Supervis by Sr. Tech. &Staff	Functional	-----
----	---	-----------	---	----------------------------	-------------------------	---------	--	------------	-------

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 04/02/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
4.	Bowa Cautery a) S.N. 3500/336 b) S.N. 3500/681	Bowa Germany	Life Aid Device 9968370861	Tr. taken from E.O.T. 17/6/08 (681) &17/9/08 (336)	EURO 81,611/- each	Monthly	a) S.N. 3500/336 (04/03/2014) by Daleep from Rohit Surgical b) S.N. 3500/681	S.N. 3500/336 Equipment is working but neutral electrode socket need to be changed S.N. 3500/681	Cautery Machine is working Work is on process

							(26/05/14) by Daleep from Rohit Surgical Supervis by Mr. Tech. &Staff	hand activation is not working	
--	--	--	--	--	--	--	---	-----------------------------------	--

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 04/02/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
--------	--------------------	-------------------	------------------	------------------	---------------------	---	---	---------------------------	---

5.	Ceiling Light (Three)	Berch told GMBH Germany	Rohit Surgical Pvt. Ltd. 9311088110	Rd. from store 06/05/08	EURO 19,396.5/- per unit	Monthly	03/06/14 by Mr. Daleep from Rohti Surgical Supervis by Sr. Tech. &Staff	(Neuro O.T.) One Dome not working due to faulty control and power supply	Work on process from M &R
6.	O.T. Table Surgery T.S. 200 Takuchi S.N. 0080	Takuchi	Rohanika Elec. &Med. System 9212791757	Rd. from store 10/06/08	Yen 22,61,070/-	Monthly	28/01/15 by Mr. Bhavesh Supervis by Sr. Tech. &Staff	Table is working Only Right leg plate locking lever is not working properly needs to be replaced Estimate will be provided by Comp. Eng.	-----
7.	O.T. Table Surgery (Surginox) S.no. 15068	Steris Surginox USA	Steris India Pvt. Ltd.	Rd. on 16/05/2012	\$ 15220/-	Monthly	17/12/14 by Mr. Amamul Supervis by Sr. Tech. &Staff	Functional	-----

**Dr. Ram Manohar Lohia Hospital
New Delhi**

Detail of Patient Care Equipment Equipments Costing more than Rs. 1.00 lakh (Purchased on or after 01.04.2007)

Department : Trauma O.T. 5th Floor(Surgery)

Dated :- 04/02/15

S. No.	Equipments Name of	Name Manufacturer	Name of Supplier	Date of Purchase	Cost of Procurement	Recommended Frequency for Maintenance (Yearly, Half Yearly, Monthly, Quarterly)	Date of Last Maintenance Check, Name of Supervising	Whether Functional or Not	If Not Functional details of Action taken
8.	Ventilator 1). S.N. 3121118350 2). S.N. 3121123870 3). S.N. 3121124100 4). S.N. 3121124370	Covidien	Medi Works A-56, IIIrd floor, Lajpat Nagar-II, New Delhi. Ph : 91-11- 46563900, 93105 66668	Rd. from store 12/08/2013	Cost @ U.S. \$ 19,200/- each	Monthly	28/01/15 by Mr. Kumar Keshav Supervis by Sr. Tech. &Staff	Functional	-----
9.	Post op. ICU Beds (04) 1. S.no. L231 HE 3606 2. S.no. M118 HE 1414 3. S.no. M113 HE 1388 4. S.no. M113 HE 1387 Model – 405 Electric Bed	Hill Rom	Rohanika Electronics &Medical Systems, New Delhi	Rd. on 05/01/2012	Rs. 1,47,232/-	Monthly	21/01/2015 by Mr. Ashok Kumar Sagar Supervised by Sr. Tech. &Staff	Functional	-----

