

**GOVERNMENT OF INDIA**  
**DR. RAM MANOHAR LOHIA HOSPITAL, NEW DELHI**

Tender No. 4-38/2019-LH(S)(ABVIMS) 2960.

Dated: 24/5/2021

[Please quote this reference at the top of sealed envelope]

(Single Bid)

**Limited Tender for the procurement of Lypocheck clinical chemistry control 1 & 2 and Liquicheck spinal fluid control 1 & 2 (Specification enclosed) (Specification enclosed)**

Date of Online Closing the Tender: 12.06.2021

Time of Closing the Tender: 5.00 PM

Date of Online Opening the Tender: 14.06.2021

Time of Opening the Tender: 2.30 PM

**Sub: Conclusion rate agreement for the supply of items, mentioned in the enclosed required for the use of Dr. RML Hospital, New Delhi.**

It is proposed to enter into a rate agreement between your firm herein after referred to as contractor and the Medical Superintendent, Dr. RML Hospital for and on behalf of the President of India, herein after referred to as purchaser for the supply of items listed in the enclosed schedule.

1. **The agreement shall be valid for two year from the date of finalization of tender.** However the Medical Superintendent reserve the right at his discretion to increase or decrease the validity period, after due date of expiry, depending on essentiality of the item and position of new tender agreement.
2. If the tenderer fails to **supply the items within 30 days of dispatch of supply order for Indian items & 90 days for imported items, liquidated damages for a sum equivalent to 0.5% of the value of the order for each week of delay or part thereof up to a maximum of 10% will be deducted from the contract price and risk purchase will be made from next lowest at the cost of supplier without any notice.** In the event of failure to supply the items within delivery period, in addition to recovery already stated, difference of amount will be charged from defaulter firm and to be **deducted from the earnest money/any other dues.** However in case of non-supply, the earnest money will be forfeited. However, Medical Superintendent has the discretion to extend the delivery period without penalty if satisfied by the reasons of delay.
3. It will be the condition that the prices charged for the stores supplied to the hospital shall in no event exceed the lowest, at which the contractor sells the stores to any other person/institution. If any time during the contract period the contractor reduces the sale price entered into the rate agreement, he will notify such reduction to the Medical Superintendent.
4. **Rates should be quoted on F.O.R. (Dr. R.M.L. Hospital) on INR basis only. NMIC & CDEC will not provided by the hospital for items quoted in Indian currency.**
5. The quantity shown in the schedule is an estimated requirement and there is no guarantee that the order for the same will be placed, it may change as per requirement of the hospital or can be totally cancelled after giving order by competent authority.
6. No figures or words should be overwritten. In-correct words/figures should be scored off and re-written under your initials.
7. **Excise duty, GST and other taxes, where legally livable should be distinctly shown along with price quoted, in case it is not done, no claim regarding excise duty/sales tax and GST will be entertained in future.** In case there is a revision in the existing rate of duty by the Government, the same will be paid subject to production of satisfactory documentary proof.
8. Quotations should be as per Sl. No. of tender and strictly according to specifications mentioned in the schedule.
9. Decision of the Medical Superintendent will be final & binding to all concerned.
10. All the damages in transit or manufacturing defect etc. detected at the time of supply shall be at the risk & cost of contractor.

11. All samples/supplies rejected shall be removed by the supplier or through his authorized representatives within the period of 15 days from the date of submission/ intimation. If the goods are not collected within the stipulated period it will be disposed off in any manner deemed fit by the purchaser.
12. The Medical Superintendent, Dr. R.M.L. Hospital reserves the right at his discretion to accept or reject any tender at any stage in part or full without assigning any reason whatsoever.
13. All stores received shall be subject to inspection on receipt of the stores & the purchaser shall have full discretion to accept or reject either the entire supply or any part thereof and his decision in this regard shall be final & binding on the supplier.
14. **This tender together with the schedule shall be deposited ONLINE ON CPPP PORTAL in two parts i.e., Technical Bid DOCUMENTS IN PDF FORMAT & FILLED Price Bid in excel format ,which is available in online tender at CPP portal.**
15. **Latest ITCC & STCC should be enclosed in the Technical bid.**
16. **The offer's of manufacturer or their authorized distributor/dealer will only be considered. The name of manufacturer should be mentioned against each item. Latest Authorization certificate should be attached with quotation, failing which their tender will be rejected.**
17. **The tenderer should strictly quote appropriate product of a single manufacturer. Offer of multiple manufacturer against a particular item will not be considered.**
18. **Tenderer should submit the latest performance certificate for last two years from any Government Institutes along with tender documents.**
19. As per O.M no. Z-28015/34/20117-EPW dated 08<sup>th</sup> August 2018 it is advised that as per Rule-149 of GFR, the procurement of goods or services by Ministries or Departments is mandatory for goods available on GeM portal. Hospital reserve the right to procure through tender or from other sources as per the instructions of Government of India.
20. **Buyers cannot be forced to buy any items through tender, if it is available on GeM Portal during validity of tender.**
21. **Preference to Make in India:** As per the order issued by Department of Industrial Policy and Promotion (DIPP) vide order no. P-45021/2/2017-BE-II dated 15.06.2017 and subsequent orders ,the purchaser reserves the right to give preference to the local supplier.
22. **As per office memorandum no. F.9/4/2020-PPD Government of India, Ministry of Finance Department of Expenditure, Procurement policy Division dated 12<sup>th</sup> November 2020, No Earnest Money is required and bidders may be asked to submit a Bid Security Declaration . Accordingly a declaration on bidders letter head may be submitted by bidder stating that if bidder withdraw or modify their bids during period of validity , they will be suspended for the one year and will not be eligible to bid for ABVIMS & Dr. RML Hospital, New Delhi tender from the date of suspension.**
23. Tenderer should submit an undertaking that all the terms and condition as mentioned in this tender are acceptable to them, along with tender document.
24. Failure to supply any item quoted, after receiving the order , will be treated as a disqualification for taking part in future tender.
25. The Medical Superintendent reserves the right to impose any penalty deemed appropriate if supplied material is found defective and not as per samples submitted, including not placing any further orders for the concerned item.
26. Training should be provided at the installation site at their own expenses.



**C.M.O I/c (Procurement)  
For: Medical Superintendent**

### **Specification for Clinical Chemistry Control**

1. The quality control material should be supplied in the lyophilized form.
2. No special diluents should be required for reconstitution of the lyophilized control material.
3. The control material should be human serum based.
4. The control material should be stable at least for 30 days after reconstitution at -20 degree C.
5. The stability of the parameters should be for at least 7 days after reconstitution at 2-8 degree C.
6. The range for the following parameters should be present in the quoted quality control material- glucose, Urea, Creatinine, Uric Acid, T.bil, D. Bil, SGOT, SGPT, ALP, Protein(Total), Albumin, Cholesterol(total), Direct HDL, Triglycerides, Sodium , Potassium, Chloride, Calcium(total), Phosphorus, CreatinineKinase, LDHl, Iron, TIBC, UIBC, Amylase, Lipase, Magnesium, Lactate, Carbamazepine, Phenyton, Valproic Acid, Lithium.
7. The control range should be present for the instruments like vitros, Olympus and Transasia in the quality control material as the department has all these equipments at present in different labs.
8. The control material should have two different levels of biochemical parameters.

### **Specification for cerebrospinal fluid controls**

1. Control should be a multi-analyte control with relevant concentrations for Cerebro spinal fluid tests especially micro-protein by the pyragallol red method.
2. Control should be Lyophilised or ready to use.
3. Control should be made from human based material.
4. Control should have adequate ( $\geq 1.5$  years) stability at 2-8 degree centigrade.
5. Control should have 2-4 weeks open vial (punctured vial) stability at 2-8 degree centigrade.
6. Should have fill size of 2.5 – 3 ml.
7. Should incorporate analytes like micro protein, glucose, chloride, Albumin, protein electrophoresis , immune globulins , sodium, lactate, lactate dehydrogenase(optional).
8. Certification: Should have quality certification like FDA/CE marking.

