



दूरभाष रू23365525
Tele No. : 23365525



डा. रा.म.लो.अ. 27(अ)
D.R.M.L.H. - 27 (A)

भारत सरकार / GOVERNMENT OF INDIA
डा. राम मनोहर लोहिया अस्पताल, नई दिल्ली
DR. RAM MANOHAR LOHIA HOSPITAL, NEW DELHI
(PROCUREMENT SECTION)

File No. Proc-11012/128/2024-Proc. Section-Dr. RMLH / 789
E- 22716

NEW DELHI Dated 14/2/25

CORRIGENDUM

Following amendments in online Bidding No.GEM/2024/B/5889091 dated 01.02.2025 for Group Bid for Procurement of Emergency Medicine Department on Turnkey is hereby authorized:-

Sl. No	Type	Existing Technical Specification	Amended Technical Specification To be read as-
1	Technical Specification of High End ICU Ventilator	Point No. 17 - Should have Battery backup with charging system - 30 minutes	Point No. 17 - Should have Battery backup with charging system - 30 minutes or more
		Accessories Point No.4 Flow sensors should be at the distal end Reusable - 02 Nos. and Disposable - 10 Nos.	Accessories Point No.4 Flow sensors should be at the distal end Reusable - 10 Nos. and Disposable - 02 Nos.
3	Technical Specification of Syringe Infusion Pump	Point No. 1 - Flow rate range should be 0.1 ml/hr to 1200 ml/h with accuracy of $\pm 1\%$	Point No. 1 - Flow rate range should be 0.1 ml/hr to 999 ml/h with accuracy of $\pm 1\%$
		Point No. 5 - Should have drug library of minimum 3800 drugs categorized in user defined 19 categories with facility to set all infusion parameters like soft limit, hard limit, bolus dose etc	Point No. 5 :- Should have drug library of minimum 3000 drugs or more categorized in user defined 15 categories or more with facility to set all infusion parameters like soft limit, hard limit, bolus dose etc
		Point No. 7 - Should have facility to upload drug library simultaneously through a single interface in the station with up to 4 infusion pumps in system with an external hardware	Point No. 7 - Should have a facility to upload drug library simultaneously through a single interface in the station with up to 03 or more infusion pumps in system with an external hardware
		Point No. 9 -Should have settable KVO option ranging from 0.1 to 5ml/hr with feature to keep it off	Point No. 9 -Should have settable KVO option ranging from 0.1 to 03 or more ml/hr with feature to keep it off
		Point No. 18 - Should save 1500 data log events in real time and should have graphical history of Volume/ dose infused, pressure and flow rate	Point No. 18 - Should save 1000 or more data log events in real time and should have graphical history of Volume/ dose infused, pressure and flow rate
		Additional Specification of Syringe Infusion Pump Point No. 1 - One Docking stand should be supplied for every 4 syringe Pump. The product should be min. IP22 certified.	Additional Specification of Syringe Infusion Pump Point No. 1 - One Docking stand should be supplied for every 03 or more syringe Pump . The product should be min. IP22 certified.

All other terms and conditions of the subject tender enquiry shall remain unchanged.

OFFICER I/C PROCUREMENT
For Medical Superintendent